



TDS CONDUCTIVITY METER



Model No.-
UPC-TDM-204

UPC Instruments Pvt. Ltd.

An ISO 9001:2015, 14001-2015 CE & TÜV SÜD Certified Company

TDS CONDUCTIVITY METER "UPC-TDM-204"

Make: "UPC"

TDS /Conductivity Range:- 0-20000ppm

Resolution :-0.1ppm

Auxiliary Power :- 10A/230V



ENCLOSURE:

Mounting : Panel Mounting

Position : Panel

Enclosure : 96(W)X96(L)X53(D)mm

Cut out : 92(L)mm X92(W)mm.

Terminal : Screw On Terminals.

KEYS:

SET : This key is used to change the controller from RUN mode to
FUNCTION Mode for parameter setting

INC : UP scroll the parameter & parameter value.

DEC : Down scroll the parameter & parameter value.

LED INDICATIONS:

Relay ON	: Green Led
Glow	: Relay ON
Not Glowing	: Relay Off
Relay Output	: NO-NC
TDS	: Red Led Glow - TDS Selected
Not Glowing	: TDS not Selected
Conductivity	: Red LED
Glow	: Conductivity Selected
Not Glowing	: Conductivity not selected
Supply Voltage	: 230VAC 10% 50Hz.

WARRANTY

All meter are warranted against any manufacturing defect for a period of 12 months from date of supply, provided the meter has not been misused, damaged, installed for services it is not recommended or the seal has been tampered with. The company shall be liable to furnish part/ parts thereof or full water meter as the company may deem fit.

UPC INSTRUMENTS PVT.LTD.
AN ISO 9001:2015, 14001-2015 CE & TÜV SÜD CERTIFIED COMPANY

Office: Plot No. 18, Towel Market, Gohana Road, Panipat-132103 (HARYANA)

Manufacturing Unit: Plot no.125,Near Tehsil, Dinger Majra Road, Gharaunda, Karnal-132114 (HARYANA)

✉ : info@unitechmeter.com ✉ : ho@unitechmeter.com ✎ : www.unitechmeter.com

Due to continuous development program, design and data in this leaflet are subject to change without prior notice.